
AccuSpike-IR Product Specification and Analysis Report

Product: 1.5 mL polypropylene vial containing 100 each, counted by flow-cytometry cell sorting, of *Giardia lamblia*, H3 isolate, cysts and *Cryptosporidium parvum*, Iowa isolate, oocysts suspended in 0.75 mL of a de-ionized water/0.01% Tween 20 solution. AccuSpike-IR is designed for percent recovery determination with matrix (environmental) and reagent water samples by US EPA Methods 1622 and 1623.

Species/genus identification method: Direct immunofluorescence microscopy with genus-specific monoclonal antibodies; also phase microscopy.

Purification method: Cysts and oocysts are purified from feces by sucrose and Percoll density gradient centrifugation.

Quantitation method: Cells enumerated using flow cytometry.

Storage Conditions: 4 C. DO NOT FREEZE.

Inactivation/Sterilization: Gamma irradiation.

Lot# 44**Preparation: 08/13/09****Expiration: 11/05/09 for *Giardia* & *Cryptosporidium* analysis****Expiration: 12/03/09 for *Cryptosporidium* analysis ONLY**

Calibration Data

Test:	Spike Preparation	Spike Preparation
	<i>Cryptosporidium</i>	<i>Giardia</i>
Organism:	<i>Cryptosporidium</i>	<i>Giardia</i>
Source of Organism:	Experimentally infected mice.	Experimentally infected gerbils.
Organism strain:	CpAZ	H3
Stock suspension lot number:	80309-23	090805
Date cells collected by source:	8/3/09	8/5/09
Date of initial calibration:	8/13/09	8/13/09
Age of cells (in days):	10	8
Storage media:	De-ionized Water/0.01% Tween 20	De-ionized Water/0.01% Tween 20
Storage temperature:	4 C	4 C
Viability (PI) before Irradiation	97.6%	98.1%
Mean of the counts:	98.61	99.83
Standard deviation of the counts:	1.80	1.87
Relative standard deviation:	1.83	1.88

Notes from Sorting Facility:

(1) Mean, standard deviation and relative standard deviation are calculated from a minimum of 12 calibration verification samples per set of 10 standards.

(2) Parasites are evaluated for general quality and intactness under DIC microscopy prior to use.